

SUPERVISOR TEST

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 01/11/2021

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:09
INTERNAL STANDARD	VERIFIED	00:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:11
BLANK TEST	0.00	00:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:13

