

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 02/03/2021

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:25
INTERNAL STANDARD	VERIFIED	22:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:26
BLANK TEST	0.00	22:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:27
BLANK TEST	0.00	22:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:28
BLANK TEST	0.00	22:29

