

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 01/30/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:06
INTERNAL STANDARD	VERIFIED	14:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:06
BLANK TEST	0.00	14:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:07
BLANK TEST	0.00	14:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:09
BLANK TEST	0.00	14:10

