

SOLUTION CHANGE

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 02/06/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:27
INTERNAL STANDARD	VERIFIED	23:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:28
BLANK TEST	0.00	23:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:29
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:30
BLANK TEST	0.00	23:31

