

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 02/08/2021

OPERATOR NAME:
CORY L HURLBURT
Permit #: 37017
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:07
INTERNAL STANDARD	VERIFIED	17:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:07
BLANK TEST	0.00	17:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:09
BLANK TEST	0.00	17:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:10
BLANK TEST	0.00	17:11

H 757
#708

