

SOLUTION CHANGE

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 02/01/2021

OPERATOR NAME:  
CHRISTOPHER J MURRAY  
Permit #: 46226  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:02
INTERNAL STANDARD	VERIFIED	02:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:02
BLANK TEST	0.00	02:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:04
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:05
BLANK TEST	0.00	02:06

CM #821

