

SOLUTION CHANGE

 CAMBRIDGE VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 126706

Date: 03/01/2021

OPERATOR NAME:
 ROBERT M DANKO
 Permit #: 39105
 Operator Agency/Dept: CAMBRIDGE VILL. PD
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:25
INTERNAL STANDARD	VERIFIED	08:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:25
BLANK TEST	0.00	08:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:26
BLANK TEST	0.00	08:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:28
BLANK TEST	0.00	08:28

