

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 02/21/2021

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:03
INTERNAL STANDARD	VERIFIED	23:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:04
BLANK TEST	0.00	23:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:05
BLANK TEST	0.00	23:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:06
BLANK TEST	0.00	23:07

