

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 02/23/2021

OPERATOR NAME:  
RICHARD J ARNOLD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:10
INTERNAL STANDARD	VERIFIED	23:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:10
BLANK TEST	0.00	23:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:11
BLANK TEST	0.00	23:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:13
BLANK TEST	0.00	23:14

