

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 03/03/2021

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:09
INTERNAL STANDARD	VERIFIED	21:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:10
BLANK TEST	0.00	21:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:11
BLANK TEST	0.00	21:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:12
BLANK TEST	0.00	21:13

