

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 03/07/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:23
INTERNAL STANDARD	VERIFIED	04:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:24
BLANK TEST	0.00	04:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:25
BLANK TEST	0.00	04:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:26
BLANK TEST	0.00	04:27

