

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 03/09/2021

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:53
INTERNAL STANDARD	VERIFIED	19:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:54
BLANK TEST	0.00	19:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:55
BLANK TEST	0.00	19:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:56
BLANK TEST	0.00	19:57

