

SOLUTION CHANGE

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 03/06/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:32
INTERNAL STANDARD	VERIFIED	17:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:33
BLANK TEST	0.00	17:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:34
BLANK TEST	0.00	17:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:35
BLANK TEST	0.00	17:36

