

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 03/13/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:08
INTERNAL STANDARD	VERIFIED	18:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:08
BLANK TEST	0.00	18:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:09
BLANK TEST	0.00	18:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:11
BLANK TEST	0.00	18:11

