

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 03/13/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:35
INTERNAL STANDARD	VERIFIED	06:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:36
BLANK TEST	0.00	06:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:37
BLANK TEST	0.00	06:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:38
BLANK TEST	0.00	06:39

