

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 03/17/2021

OPERATOR NAME:  
RICHARD J ARNOLD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:43
INTERNAL STANDARD	VERIFIED	18:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:44
BLANK TEST	0.00	18:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:45
BLANK TEST	0.00	18:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:46
BLANK TEST	0.00	18:47

