

SOLUTION CHANGE

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 03/18/2021

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:13
INTERNAL STANDARD	VERIFIED	16:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:14
BLANK TEST	0.00	16:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:15
BLANK TEST	0.00	16:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:17
BLANK TEST	0.00	16:17

