

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 03/23/2021

OPERATOR NAME:
 RICHARD J ARNOLD
 Permit #: 33860
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20500

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 20:01 |
| INTERNAL STANDARD | VERIFIED | 20:02 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 20:02 |
| BLANK TEST | 0.00 | 20:03 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 20:03 |
| BLANK TEST | 0.00 | 20:04 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 20:04 |
| BLANK TEST | 0.00 | 20:05 |

