

SUPERVISOR TEST

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 03/28/2021

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:18
INTERNAL STANDARD	VERIFIED	03:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:19
BLANK TEST	0.00	03:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:20
BLANK TEST	0.00	03:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:21
BLANK TEST	0.00	03:22

