

SUPERVISOR TEST


WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 03/27/2021

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:08
INTERNAL STANDARD	VERIFIED	14:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:09
BLANK TEST	0.00	14:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:10
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:12



SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dnt: 113806

Date: 03/27/2021

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:08
INTERNAL STANDARD	VERIFIED	14:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:09
BLANK TEST	0.00	14:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:10
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:12

