

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 04/01/2021

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:11 |
| INTERNAL STANDARD | VERIFIED | 00:12 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 00:12 |
| BLANK TEST | 0.00 | 00:13 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 00:13 |
| BLANK TEST | 0.00 | 00:14 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 00:14 |
| BLANK TEST | 0.00 | 00:15 |

