

SUPERVISOR TEST

-----  
 FORT EDWARD  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 115708  
 -----

Date: 03/21/2021

OPERATOR NAME:  
 SGT DEAN E WATKINS  
 Permit #: 39523  
 Operator Agency/Dept: FORT EDWARD VILL. PD  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:11
INTERNAL STANDARD	VERIFIED	23:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:11
BLANK TEST	0.00	23:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:13
BLANK TEST	0.00	23:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:14
BLANK TEST	0.00	23:15

