

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 04/16/2021

OPERATOR NAME:  
BRAD M LYON  
Permit #: 44978  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:33
INTERNAL STANDARD	VERIFIED	10:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:33
BLANK TEST	0.00	10:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:35
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:36
BLANK TEST	0.00	10:37

