

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 04/12/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:13
INTERNAL STANDARD	VERIFIED	01:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:13
BLANK TEST	0.00	01:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:14
BLANK TEST	0.00	01:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:16
BLANK TEST	0.00	01:16

