

SUPERVISOR TEST

WASHINGTON COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106406

Date: 03/17/2021

OPERATOR NAME:  
 CARLEIGH B BATCHELDER  
 Permit #: 48533  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:48
INTERNAL STANDARD	VERIFIED	19:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:48
BLANK TEST	0.00	19:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:50
BLANK TEST	0.00	19:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:51
BLANK TEST	0.00	19:52

*CB*  
*# 708*

