

SOLUTION CHANGE

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 04/01/2021

OPERATOR NAME:  
CORY L HURLBURT  
Permit #: 37017  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21020

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 06:41 |
| INTERNAL STANDARD       | VERIFIED | 06:41 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 06:41 |
| BLANK TEST              | 0.00     | 06:42 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 06:43 |
| BLANK TEST              | 0.00     | 06:43 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 06:44 |
| BLANK TEST              | 0.00     | 06:45 |

CA  
#757

