

SUPERVISOR TEST

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 04/26/2021

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:20
INTERNAL STANDARD	VERIFIED	07:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:24

