

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 04/24/2021

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:15
INTERNAL STANDARD	VERIFIED	22:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:16
BLANK TEST	0.00	22:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:17
BLANK TEST	0.00	22:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:18
BLANK TEST	0.00	22:19

