

SOLUTION CHANGE

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 05/01/2021

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:42
INTERNAL STANDARD	VERIFIED	18:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:42
BLANK TEST	0.00	18:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:44
BLANK TEST	0.00	18:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:45
BLANK TEST	0.00	18:45

