

SUPERVISOR TEST

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 04/28/2021

OPERATOR NAME:  
 RICHARD J ARNOLD  
 Permit #: 33860  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:06
INTERNAL STANDARD	VERIFIED	19:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:06
BLANK TEST	0.00	19:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:07
BLANK TEST	0.00	19:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:09
BLANK TEST	0.00	19:10

