

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 04/17/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:13
INTERNAL STANDARD	VERIFIED	14:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:14
BLANK TEST	0.00	14:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:15
BLANK TEST	0.00	14:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:17

