

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 04/19/2021

OPERATOR NAME:
WILLIAM M GREENE
Permit #: 48760
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:21
INTERNAL STANDARD	VERIFIED	23:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:22
BLANK TEST	0.00	23:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:23
BLANK TEST	0.00	23:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:24
BLANK TEST	0.00	23:25

WB
#647

