

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 04/20/2021

OPERATOR NAME:  
CHRISTOPHER J DAVIDSEN  
Permit #: 40077  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:10
INTERNAL STANDARD	VERIFIED	16:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:10
BLANK TEST	0.00	16:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:11
BLANK TEST	0.00	16:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:12
BLANK TEST	0.00	16:13

