

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 05/01/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:46
INTERNAL STANDARD	VERIFIED	17:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:47
BLANK TEST	0.00	17:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:48
BLANK TEST	0.00	17:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:49
BLANK TEST	0.00	17:50

