

SUPERVISOR TEST

WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 05/03/2021

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:30
INTERNAL STANDARD	VERIFIED	06:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:30
BLANK TEST	0.00	06:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:31
BLANK TEST	0.00	06:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:32
BLANK TEST	0.00	06:33

