

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 04/11/2021

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:32
INTERNAL STANDARD	VERIFIED	23:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:33
BLANK TEST	0.00	23:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:34
BLANK TEST	0.00	23:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:35
BLANK TEST	0.00	23:36

