

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 04/18/2021

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:08
INTERNAL STANDARD	VERIFIED	23:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:08
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:10
BLANK TEST	0.00	23:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:11
BLANK TEST	0.00	23:12

