

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 05/10/2021

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:19
INTERNAL STANDARD	VERIFIED	00:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:19
BLANK TEST	0.00	00:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:20
BLANK TEST	0.00	00:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:22
BLANK TEST	0.00	00:23

