

SOLUTION CHANGE

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 05/11/2021

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:50
INTERNAL STANDARD	VERIFIED	22:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:50
BLANK TEST	0.00	22:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:52
BLANK TEST	0.00	22:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:53
BLANK TEST	0.00	22:54

