

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 05/04/2021

OPERATOR NAME:
WILLIAM M GREENE
Permit #: 48760
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:55
INTERNAL STANDARD	VERIFIED	13:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:56
BLANK TEST	0.00	13:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:58
BLANK TEST	0.00	13:59

