

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 05/10/2021

OPERATOR NAME:  
KATHERINE M PADDOCK  
Permit #: 45899  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:18
INTERNAL STANDARD	VERIFIED	14:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:19
BLANK TEST	0.00	14:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:20
BLANK TEST	0.00	14:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:21
BLANK TEST	0.00	14:22

*KPW*

