

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 05/17/2021

OPERATOR NAME:  
ETHAN J MACURA  
Permit #: 44080  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:06
INTERNAL STANDARD	VERIFIED	20:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:06
BLANK TEST	0.00	20:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:07
BLANK TEST	0.00	20:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:09
BLANK TEST	0.00	20:10

*Eth*  
#624

