

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 05/24/2021

OPERATOR NAME:
MATHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:48
INTERNAL STANDARD	VERIFIED	06:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:50
BLANK TEST	0.00	06:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:52

