

SUPERVISOR TEST

WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 05/16/2021

OPERATOR NAME:
 BRYAN J GRECO
 Permit #: 44077
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:48
INTERNAL STANDARD	VERIFIED	01:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:49
BLANK TEST	0.00	01:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:51
BLANK TEST	0.00	01:52

