

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 06/07/2021

OPERATOR NAME:  
CARLEIGH B BATCHELDER  
Permit #: 48533  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:54
INTERNAL STANDARD	VERIFIED	23:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:55
BLANK TEST	0.00	23:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:56
BLANK TEST	0.00	23:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:57
BLANK TEST	0.00	23:58

CBB  
#708

