

SOLUTION CHANGE

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 05/31/2021

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:43
INTERNAL STANDARD	VERIFIED	15:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:43
BLANK TEST	0.00	15:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:44
BLANK TEST	0.00	15:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:46
BLANK TEST	0.00	15:47

