

SUPERVISOR TEST

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 06/07/2021

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:18
INTERNAL STANDARD	VERIFIED	11:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:19
BLANK TEST	0.00	11:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:20
BLANK TEST	0.00	11:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:21
BLANK TEST	0.00	11:22

