

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 05/22/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:09
INTERNAL STANDARD	VERIFIED	12:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:09
BLANK TEST	0.00	12:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:11
BLANK TEST	0.00	12:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:12
BLANK TEST	0.00	12:13

