

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 05/29/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:42
INTERNAL STANDARD	VERIFIED	06:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:43
BLANK TEST	0.00	06:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:44
BLANK TEST	0.00	06:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:45
BLANK TEST	0.00	06:46

