

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 05/29/2021

OPERATOR NAME:
 RYAN M WOODWARD
 Permit #: 41955
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:26
INTERNAL STANDARD	VERIFIED	00:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:26
BLANK TEST	0.00	00:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:27
BLANK TEST	0.00	00:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:29
BLANK TEST	0.00	00:29

